

Hillrom

achieves the desired results. This may require validation and routine monitoring of your facility's actual process. This also may require validation of your actual process.
NOTE The main handle and bottom cap components of handpieces marked "AUTOCLEAN" are compatible with autoclaves and sterilizers which are pre-identified for facilities who wish to perform either method after cleaning and intermediate-level disinfection.

Cleaning and intermediate-level disinfection

- Separate blade assembly from handle and place handle into suitable container for subsequent reprocessing. (See Figure 1. Do not place handle with sharp edges into container.)
- Prevent the handle from drying (i.e., wrap/cover in moist germicidal wipe).

Preparation for decontamination

Select a cleaning/sterilization/autoclave/autoclave/autoclave/autoclave-based germicidal cleaner labeled suitable for use on healthcare equipment and capable of removing all organic residues. EPA-registered disinfectants: https://www.epa.gov/pesticide/registered/selected-epa-registered-disinfectants. Outside of the U.S., consult applicable regulatory agency for the equivalent quality of an ammonium isopropanol germicidal cleaner.

Remove batteries. See Figure 2.
Cleaning and intermediate-level disinfection
Follow the germicidal wipe manufacturer's instructions to clean all exposed surfaces of the main handle and end cap.
CAUTION Do not immerse/soak handle and re-wipe to prevent re-exposure to excessive visible soil.
CAUTION Do not immerse/soak re-wipe to wet all surfaces and allow adequate contact time for disinfection as directed by the germicidal wipe manufacturer.
CAUTION Only use quaternary ammonium isopropanol-based germicidal wipes.

Drying

Allow components to air dry.

Maintenance, inspection, and testing

- Inspect each component area for damage or deterioration. Inspect (1) main handle REF number, (2) main handle, (3) batteries, and (4) bottom cap. See Figure 2.

WARNING Discard any component that shows evidence of damage or deterioration. Contact Hillrom Technical Support for component replacement.

Reassemble handle with new or batteries in known good condition. See Figure 2.

Attach handle to a clean and disinfected test blade in known working condition. See Figure 2.

- Blade assembly engages and locks onto handle.
- Blade assembly deploys into its locked position on handle AND lamp illuminates.
- Light output is satisfactory.

If the lamp fails to light or output is low, check or replace the lamp and/or battery.

Storage

Store handle per facility practice to allow device to remain clean, dry, and ready for service.

End of reprocessing instructions for intermediate-level cleaning.

Autoclave

NOTE The main handle and bottom cap components of handles marked "AUTOCLEAN" are compatible with the autoclave methods identified which are provided for facilities who wish to autoclave after cleaning and intermediate-level disinfection.

Disassembly

- Remove batteries and set aside. See Figure 2.
- Remove battery cover and inspect ONE of the following autoclave methods below for the main handle and bottom cap only):
 - Remove top cap and place in container for reprocessing.
 - Remove top cap, manufacturer and facility procedures in the setup and operation of autoclave equipment. Gravity autoclave settings are as follows:
 - Temperature: 132 °C (270 °F)
 - Exposure time: 3 minutes (unwrapped)
 - Minimum dry time: 1 minute
 - Pre-vacuum autoclave. Follow equipment manufacturer and facility procedures in the setup and operation of autoclave equipment. Pre-vacuum autoclave settings are as follows:
 - Temperature: 132 °C (270 °F)
 - Exposure time: 3 minutes (unwrapped)
 - Minimum dry time: 1 minute

Maintenance, inspection, and testing

- Inspect each component area for damage or deterioration. See Figure 3.

WARNING Discard any component that shows evidence of damage or deterioration. Contact Hillrom Technical Support for component replacement.

Reassemble with new or batteries in known good condition. See Figure 4.

Attach handle to a clean and disinfected test blade assembly in known working condition. See Figure 4.

- Blade assembly engages and locks onto handle.
- Blade assembly deploys into its locked position on handle AND lamp illuminates.
- Light output is satisfactory.

If the lamp fails to light or output is low, check or replace the lamp and/or battery.

Storage

Store handle per facility practice to allow device to stay clean, dry, and ready for service.

Maintenance

- Insert batteries and reinstall bottom cap. See Figure 4.

NOTE All alkaline batteries are supplied with your handle for maximum performance and are recommended as replacements; however, carbon-zinc batteries may also be used.

- Large handle, REF 60300 uses two "D" size
- Medium handle, REF 60300 uses two "C" size
- Penlight handle, REF 60400 uses two "AA" size
- Quabty handles, REF 60205 uses two "AAA" size

- Reprocess reassembled assembly as appropriate per these instructions.

Disposal

Users must adhere to all federal, state, regional, and/or local laws and regulations as they pertain to the safe disposal of medical devices and their components. If in doubt, the user of the device should first contact Hillrom Technical Support for questions regarding disposal protocols.

PATENTS/PATENT

Hillrom.com patents are owned by one or more parents. See show Internet address. The Hill Rom companies are the proprietors of European, U.S., and other patents and pending applications.

Hillrom Technical Support

For information about any Hillrom product, contact Hillrom Technical Support at hillrom.com/en-us/service-support/enrhc.html.

Notice to users and/or patient in EU

Any serious incident which has occurred in relation to the device, should be reported to the manufacturer and the competent authority of the Member State in which the user or patient is established.

Specifications

Temperature limits

Operating 32° to 104°* (0° to 40° C)

Storage/transport 2° to 49°* (32° to 120° F)

Standards and compliance

The device complies with the following standards:

EN 12776 – European standard for disinfection and sterilization

ISO 60601-1 – International standard for safety of electrical medical equipment

Country specific standards are included in the applicable Declaration of Conformity.

LiCo code

For information about any Hillrom product, contact Hillrom Technical Support at hillrom.com/en-us/service-support/enrhc.html.

Guidance and manufacturer's declaration

Users must adhere to all federal, state, regional, and/or local laws and regulations as they pertain to the safe disposal of medical devices and their components. If in doubt, the user of the device should first contact Hillrom Technical Support for questions regarding disposal protocols.

PATENTS/PATENT

Hillrom.com patents are owned by one or more parents. See show Internet address. The Hill Rom companies are the proprietors of European, U.S., and other patents and pending applications.

Hillrom Technical Support

For information about any Hillrom product, contact Hillrom Technical Support at hillrom.com/en-us/service-support/enrhc.html.

Notice to users and/or patient in EU

Any serious incident which has occurred in relation to the device, should be reported to the manufacturer and the competent authority of the Member State in which the user or patient is established.

Specifications

Temperature limits

Operating 32° to 104°* (0° to 40° C)

Storage/transport 2° to 49°* (32° to 120° F)

Standards and compliance

The device complies with the following standards:

EN 12776 – European standard for disinfection and sterilization

ISO 60601-1 – International standard for safety of electrical medical equipment

Country specific standards are included in the applicable Declaration of Conformity.

LiCo code

For information about any Hillrom product, contact Hillrom Technical Support at hillrom.com/en-us/service-support/enrhc.html.

Guidance and manufacturer's declaration

Users must adhere to all federal, state, regional, and/or local laws and regulations as they pertain to the safe disposal of medical devices and their components. If in doubt, the user of the device should first contact Hillrom Technical Support for questions regarding disposal protocols.

PATENTS/PATENT

Hillrom.com patents are owned by one or more parents. See show Internet address. The Hill Rom companies are the proprietors of European, U.S., and other patents and pending applications.

Hillrom Technical Support

For information about any Hillrom product, contact Hillrom Technical Support at hillrom.com/en-us/service-support/enrhc.html.

Notice to users and/or patient in EU

Any serious incident which has occurred in relation to the device, should be reported to the manufacturer and the competent authority of the Member State in which the user or patient is established.

Specifications

Temperature limits

Operating 32° to 104°* (0° to 40° C)

Storage/transport 2° to 49°* (32° to 120° F)

Standards and compliance

The device complies with the following standards:

EN 12776 – European standard for disinfection and sterilization

ISO 60601-1 – International standard for safety of electrical medical equipment

Country specific standards are included in the applicable Declaration of Conformity.

LiCo code

For information about any Hillrom product, contact Hillrom Technical Support at hillrom.com/en-us/service-support/enrhc.html.

Guidance and manufacturer's declaration

Users must adhere to all federal, state, regional, and/or local laws and regulations as they pertain to the safe disposal of medical devices and their components. If in doubt, the user of the device should first contact Hillrom Technical Support for questions regarding disposal protocols.

PATENTS/PATENT

Hillrom.com patents are owned by one or more parents. See show Internet address. The Hill Rom companies are the proprietors of European, U.S., and other patents and pending applications.

Hillrom Technical Support

For information about any Hillrom product, contact Hillrom Technical Support at hillrom.com/en-us/service-support/enrhc.html.

Notice to users and/or patient in EU

Any serious incident which has occurred in relation to the device, should be reported to the manufacturer and the competent authority of the Member State in which the user or patient is established.

achieves the desired results. This may require validation and routine monitoring of your facility's actual process. This also may require validation of your actual process.
NOTE The main handle and bottom cap components of handpieces marked "AUTOCLEAN" are compatible with autoclaves and sterilizers which are pre-identified for facilities who wish to perform either method after cleaning and intermediate-level disinfection.

Cleaning and intermediate-level disinfection

- Separate blade assembly from handle and place handle into suitable container for subsequent reprocessing. (See Figure 1. Do not place handle with sharp edges into container.)
- Prevent the handle from drying (i.e., wrap/cover in moist germicidal wipe).

Preparation for decontamination

Select a cleaning/sterilization/autoclave/autoclave/autoclave/autoclave-based germicidal cleaner labeled suitable for use on healthcare equipment and capable of removing all organic residues. EPA-registered disinfectants: https://www.epa.gov/pesticide/registered/selected-epa-registered-disinfectants. Outside of the U.S., consult applicable regulatory agency for the equivalent quality of an ammonium isopropanol germicidal cleaner.

Remove batteries. See Figure 2.
Cleaning and intermediate-level disinfection
Follow the germicidal wipe manufacturer's instructions to clean all exposed surfaces of the main handle and end cap.
CAUTION Do not immerse/soak handle and re-wipe to prevent re-exposure to excessive visible soil.
CAUTION Do not immerse/soak re-wipe to wet all surfaces and allow adequate contact time for disinfection as directed by the germicidal wipe manufacturer.
CAUTION Only use quaternary ammonium isopropanol-based germicidal wipes.

Drying

Allow components to air dry.

Maintenance, inspection, and testing

- Inspect each component area for damage or deterioration. Inspect (1) main handle REF number, (2) main handle, (3) batteries, and (4) bottom cap. See Figure 2.

WARNING Discard any component that shows evidence of damage or deterioration. Contact Hillrom Technical Support for component replacement.

Reassemble handle with new or batteries in known good condition. See Figure 2.

Attach handle to a clean and disinfected test blade in known working condition. See Figure 2.

- Blade assembly engages and locks onto handle.
- Blade assembly deploys into its locked position on handle AND lamp illuminates.
- Light output is satisfactory.

If the lamp fails to light or output is low, check or replace the lamp and/or battery.

Storage

Store handle per facility practice to allow device to remain clean, dry, and ready for service.

End of reprocessing instructions for intermediate-level cleaning.

Autoclave

NOTE The main handle and bottom cap components of handles marked "AUTOCLEAN" are compatible with the autoclave methods identified which are provided for facilities who wish to autoclave after cleaning and intermediate-level disinfection.

Disassembly

- Remove batteries and set aside. See Figure 2.
- Remove battery cover and inspect ONE of the following autoclave methods below for the main handle and bottom cap only):
 - Remove top cap and place in container for reprocessing.
 - Remove top cap, manufacturer and facility procedures in the setup and operation of autoclave equipment. Gravity autoclave settings are as follows:
 - Temperature: 132 °C (270 °F)
 - Exposure time: 3 minutes (unwrapped)
 - Minimum dry time: 1 minute
 - Pre-vacuum autoclave. Follow equipment manufacturer and facility procedures in the setup and operation of autoclave equipment. Pre-vacuum autoclave settings are as follows:
 - Temperature: 132 °C (270 °F)
 - Exposure time: 3 minutes (unwrapped)
 - Minimum dry time: 1 minute

Maintenance, inspection, and testing

- Inspect each component area for damage or deterioration. See Figure 3.

WARNING Discard any component that shows evidence of damage or deterioration. Contact Hillrom Technical Support for component replacement.

Reassemble with new or batteries in known good condition. See Figure 4.

Attach handle to a clean and disinfected test blade assembly in known working condition. See Figure 4.

- Blade assembly engages and locks onto handle.
- Blade assembly deploys into its locked position on handle AND lamp illuminates.
- Light output is satisfactory.

If the lamp fails to light or output is low, check or replace the lamp and/or battery.

Storage

Store handle per facility practice to allow device to stay clean, dry, and ready for service.

Maintenance

- Insert batteries and reinstall bottom cap. See Figure 4.

NOTE All alkaline batteries are supplied with your handle for maximum performance and are recommended as replacements; however, carbon-zinc batteries may also be used.

- Large handle, REF 60300 uses two "D" size
- Medium handle, REF 60300 uses two "C" size
- Penlight handle, REF 60400 uses two "AA" size
- Quabty handles, REF 60205 uses two "AAA" size

- Reprocess reassembled assembly as appropriate per these instructions.

Disposal

Users must adhere to all federal, state, regional, and/or local laws and regulations as they pertain to the safe disposal of medical devices and their components. If in doubt, the user of the device should first contact Hillrom Technical Support for questions regarding disposal protocols.

PATENTS/PATENT

Hillrom.com patents are owned by one or more parents. See show Internet address. The Hill Rom companies are the proprietors of European, U.S., and other patents and pending applications.

Hillrom Technical Support

For information about any Hillrom product, contact Hillrom Technical Support at hillrom.com/en-us/service-support/enrhc.html.

Notice to users and/or patient in EU

Any serious incident which has occurred in relation to the device, should be reported to the manufacturer and the competent authority of the Member State in which the user or patient is established.

Specifications

Temperature limits

Operating 32° to 104°* (0° to 40° C)

Storage/transport 2° to 49°* (32° to 120° F)

Standards and compliance

The device complies with the following standards:

EN 12776 – European standard for disinfection and sterilization

ISO 60601-1 – International standard for safety of electrical medical equipment

Country specific standards are included in the applicable Declaration of Conformity.

LiCo code

For information about any Hillrom product, contact Hillrom Technical Support at hillrom.com/en-us/service-support/enrhc.html.

Guidance and manufacturer's declaration

Users must adhere to all federal, state, regional, and/or local laws and regulations as they pertain to the safe disposal of medical devices and their components. If in doubt, the user of the device should first contact Hillrom Technical Support for questions regarding disposal protocols.

PATENTS/PATENT

Hillrom.com patents are owned by one or more parents. See show Internet address. The Hill Rom companies are the proprietors of European, U.S., and other patents and pending applications.

Hillrom Technical Support

For information about any Hillrom product, contact Hillrom Technical Support at hillrom.com/en-us/service-support/enrhc.html.

Notice to users and/or patient in EU

Any serious incident which has occurred in relation to the device, should be reported to the manufacturer and the competent authority of the Member State in which the user or patient is established.

Specifications

Temperature limits

Operating 32° to 104°* (0° to 40° C)

Storage/transport 2° to 49°* (32° to 120° F)

Standards and compliance

The device complies with the following standards:

EN 12776 – European standard for disinfection and sterilization

ISO 60601-1 – International standard for safety of electrical medical equipment

Country specific standards are included in the applicable Declaration of Conformity.

LiCo code

For information about any Hillrom product, contact Hillrom Technical Support at hillrom.com/en-us/service-support/enrhc.html.

Guidance and manufacturer's declaration

Users must adhere to all federal, state, regional, and/or local laws and regulations as they pertain to the safe disposal of medical devices and their components. If in doubt, the user of the device should first contact Hillrom Technical Support for questions regarding disposal protocols.

PATENTS/PATENT

Hillrom.com patents are owned by one or more parents. See show Internet address. The Hill Rom companies are the proprietors of European, U.S., and other patents and pending applications.

Hillrom Technical Support

For information about any Hillrom product, contact Hillrom Technical Support at hillrom.com/en-us/service-support/enrhc.html.

Notice to users and/or patient in EU

Any serious incident which has occurred in relation to the device, should be reported to the manufacturer and the competent authority of the Member State in which the user or patient is established.

Specifications

Temperature limits

Operating 32° to 104°* (0° to 40° C)

Storage/transport 2° to 49°* (32° to 120° F)

Standards and compliance

The device complies with the following standards:

EN 12776 – European standard for disinfection and sterilization

ISO 60601-1 – International standard for safety of electrical medical equipment

Country specific standards are included in the applicable Declaration of Conformity.

LiCo code

For information about any Hillrom product, contact Hillrom Technical Support at hillrom.com/en-us/service-support/enrhc.html.

Guidance and manufacturer's declaration

Users must adhere to all federal, state, regional, and/or local laws and regulations as they pertain to the safe disposal of medical devices and their components. If in doubt, the user of the device should first contact Hillrom Technical Support for questions regarding disposal protocols.

PATENTS/PATENT

Hillrom.com patents are owned by one or more parents. See show Internet address. The Hill Rom companies are the proprietors of European, U.S., and other patents and pending applications.

Hillrom Technical Support

For information about any Hillrom product, contact Hillrom Technical Support at hillrom.com/en-us/service-support/enrhc.html.

Notice to users and/or patient in EU

Any serious incident which has occurred in relation to the device, should be reported to the manufacturer and the competent authority of the Member State in which the user or patient is established.

