

DECLARATION OF CONFORMITY

(in accordance with ISO/IEC 17050-1)

We declare, under our sole responsibility, that the product listed below conforms to the provisions of:

- the European Council Directive 93/42/EEC of 14 June 1993 concerning medical devices
- the Directive 2014/53/EU of the European Parliament and of the Council of 16 April 2014 on the harmonization of the laws of the Member States relating to the making available on the market of radio equipment
- the Directive 2011/65/EU of the European Parliament and of the Council of 8 June 2011 on the restriction of the use of certain hazardous substances in electrical and electronic equipment (RoHS)

Manufacturer's Name and Business Address:	Welch Allyn, Inc. 4341 State Street Road Skaneateles Falls, NY 13153 USA
EC REP	Regulatory Affairs Representative Welch Allyn Limited Navan Business Park, Dublin Road Navan, Co. Meath C15 AW22 Ireland
Product Name ^{1,3} :	ELI150C and ELI250C
REF ^{1,3}	The ELI150C and ELI250C comes in different configurations ELI150C-XXX-XXXXX ELI250C-XXX-XXXXX Where "X" can be a letter from A to Z representing the following: ELI150C - [Model] [Power Cord] [Patient Cable Leadset] - [Option1] [Option2] [Option3] [Option4] [Option5] ELI250C - [Model] [Power Cord] [Patient Cable Leadset] - [Option1] [Option2] [Option3] [Option4] [Option5]
# ^{1,3}	ELI150C 901129 – ELECTROCARDIOGRAPH ELI250C 901131 – ELECTROCARDIOGRAPH
Optional Radio equipment ² :	B&B 9373 - WLNN-AN-MR551 (DHF-THF-89515-00)
Object of the declaration ² :	802.11 a/b/g/n WLAN
Accessories and components ² :	Not Applicable

¹ applicable to the medical device directive, 93/42/EEC

² applicable to the radio equipment directive, 2014/53/EU

³ applicable to the RoHS directive, 2011/65/EU

Medical Device Conformity Assessment Route Annex ¹ :	II	
Medical Device Classification ¹ :	IIa	
Medical Device Classification Rules ¹ :	10	
GMDN Code and Term ¹ :	16231 - Electrocardiograph, professional, multichannel	
Notified Body ¹ : (CE 0459)	LNE G-MED, 1, rue Gaston Boissier 75015 Paris France EC-certificate No. 35913	
Standards Applied	Number	Title
	EN 50581 ³	Technical documentation for the assessment of electrical and electronic products with respect to the restriction of hazardous substances
	EN/IEC 62304 ¹	Medical Device Software – Software Life Cycle Processes
	EN/IEC 60601-1 ¹	Medical Electrical Equipment – Part 1: General Requirements for Basic Safety and Essential Performance
	EN/IEC 60601-1-2 ¹	Medical electrical equipment - Part 1-2: General Requirements for Safety – Collateral Standard: Electromagnetic Compatibility – Requirements and Tests
	EN/IEC 62366 ¹	Medical devices – Application of Usability Engineering to Medical Devices
	EN/IEC 60601-1-6 ¹	Medical Electrical Equipment – Part 1-6: General Requirements for Safety – Collateral Standard: Usability
	EN/IEC 60601-2-25 ¹	Medical Electrical Equipment. Part 2 25: Particular requirements for the basic safety and essential requirements of electrocardiographs
	ETSI EN 301 489 –1 ²	Electromagnetic compatibility and Radio Spectrum Matters, Part1: Common technical requirements
	ETSI EN 301 489-17	Electromagnetic compatibility and Radio Spectrum Matters

¹ applicable to the medical device directive, 93/42/EEC

² applicable to the radio equipment directive, 2014/53/EU

³ applicable to the RoHS directive, 2011/65/EU

	ETSI EN 300 328 ²	Electromagnetic compatibility and Radio Spectrum Matters, Wideband transmission systems; Data transmission equipment operating in the 2.4 GHz ISM band and using wide band modulation techniques
--	------------------------------	--

Authorised Signatory:

Mark Elliott

Mark Elliott,
Director Quality Assurance

07-OCT-2019

Date

Milwaukee, Wisconsin, USA

Place of Issue

¹ applicable to the medical device directive, 93/42/EEC

² applicable to the radio equipment directive, 2014/53/EU

³ applicable to the RoHS directive, 2011/65/EU

Document Change History

Version	Description	Author	Date
N	Initial release in new format	Marco Manduchi Steven Co Michael Lippold	2019-08-16
P	Updated EC Rep Address	Michael Lippold	2019-10-01

